

81st ARFTG Microwave Measurement Conference

(ARFTG 2013)

**Seattle, Washington, USA
7 June 2013**



IEEE Catalog Number: CFP13ARF-POD
ISBN: 978-1-4673-4981-9

Oral Session 1

Chair: David Blackham, Agilent Technologies

Cross-connected Waveguide Lines as Standards for Millimeter- and Submillimeter-wave Vector Network Analyzers 1

N. Ridler, M. Salter, NPL, Teddington, United Kingdom

New Calibration Solutions for Multi-Channel Probes using an Added Port for Thru Measurements 8

K. Jung¹, M. F. Andrews², ¹Cascade Microtech Inc., Beaverton, United States, ²Microsemi Co., Bend, United States

High Resolution Wideband Calibration Procedure for RF Time-Domain Measurement of Non-linear Devices 12

K. El-Akhdar, S. Ahmed, T. Reveyrand, G. Neveux, D. Barataud, J. Nebus, University of Limoges, XLIM UMR n°7252, Limoges Cedex, France

The Impact of Characteristic Impedance on Waveform Calibrations 16

D. F. Williams, J. Jargon, P. D. Hale, NIST, Boulder, United States

Oral Session 2

Chair: Mohamed Sayed, MMS

Comparing Accuracy of Waveguide VNA Measurement Calibrated by TRL Calibration 42 Using Different Length of Line Standard in Terahertz Band

M. Horibe, R. Kishikawa, AIST, Tsukuba, Japan

Edge-Related Calibration and Measurement Characteristics in Pulsed Profile S-parameter Analysis 29

J. Martens, Anritsu, Morgan Hill, United States

An Improved Electric Field Probe with Applications in High Efficiency PA Design and Diagnostics 53

N. Dehghan, S. Cripps, A. Porch, J. Lees, Cardiff University, Cardiff, United Kingdom

Phase-Shifterless Phase-Noise Measurement of Microwave Oscillators Using High-Q Cavity Frequency Discriminator 57

H. Gheidi¹, A. Banai², ¹University of California San Diego, La Jolla, United States, ²Sharif University of Technology, Tehran, Iran

Oral Session 3

Chair: Joe Gering, RFMD

Filter-less Diplexer Enables Active Harmonic Load-pull at Ka Band 3:

M. A. Casbon, P. J. Tasker, Cardiff University, Cardiff, United Kingdom

Next-Generation Comb Generators for Accurate Modulated Measurements 64

F. Verbeyst, M. Vanden Bossche, G. Pailloncy, National Instruments, Zaventem, Belgium

Linear Versus Nonlinear De-embedding: Experimental Investigation 68

A. Raffo¹, V. Vadalà¹, G. Avolio², G. Bosi¹, A. Nalli¹, D. Schreurs², G. Vannini¹, ¹University of Ferrara, Ferrara, Italy, ²Katholieke Universiteit Leuven, Leuven, Belgium

Synthesized Pulsed Bias for Device Characterization 73

A. Kumar Manjanna², K. Buisman¹, M. Spirito¹, M. Marchetti², M. Pelk¹, L. C. de Vreede¹, ¹Delft University of Technology, Delft, Netherlands, ²Anteverta Microwave, Delft, Netherlands

Oral Session 4

Chair: Rusty Myers, Maury Microwave

Test Port Cable Instability and VNA Measurement Errors 77

J. C. Liu, K. Wong, Agilent Technologies, Inc., Santa Rosa, United States

Figures of Uncertainty for Noise Measurements 85

M. Seelmann-Eggebert¹, B. Baldischweiler¹, B. Aja², D. Bruch¹, H. Massler¹, ¹Fraunhofer IAF, Freiburg, Germany, ²University of Cantabria, Santander, Spain

Technology for Calibrating Vector Signal Analyzer in aspect of FSK Demodulation Error 89

Z. Feng¹, Z. Rui², S. Jing-lu², S. Dan¹, S. Yuan-mao¹, G. You-gang¹, Y. Yan³, ¹Beijing University of Posts and Telecommunications, Beijing, China, ²Telecommunication Metrology Center of Ministry of Industry and Information Technology , Beijing, China, ³GuiZhou LuJun YuBeiYi BuBingShi, GuiYang, China

Interactive Forum

Chair: Thomas (Mitch) Wallis, NIST

Improvement on Technology using CW Combination to Calibrate Vector Signal Analyzer 9:

Z. Rui¹, Z. Feng¹, S. Jing-lu¹, S. Dan², S. Yuan-mao², G. You-gang², W. Gan¹, ¹Telecommunication Metrology Center of Ministry of Industry and Information Technology, Beijing, China, ²Beijing University of Posts and Telecommunications, Beijing, China

Load Pull Verification of a Novel Class B/J Design Tool: Second Harmonic Clipping Contours :7

T. Canning, P. Tasker, S. Cripps, Cardiff University, Cardiff, United Kingdom

Experiment Design for Quick Statistical FET Large Signal Model Extraction ::

I. Angelov¹, M. Ferndahl¹, M. Gavell¹, G. Avolio², D. Schreurs², ¹Chalmers University, Goteborg, Sweden, ²KU Leuven, Leuven, Belgium

Study of Self-Heating in GaAs pHEMTs using Pulsed I-V Analysis ;5

B. K. Schwitter¹, S. A. Albahrahi¹, A. E. Parker¹, L. P. Dunleavy², M. C. Heimlich¹, ¹Macquarie University, North Ryde, Australia, ²Modelithics, Inc., Tampa, United States, ³University of South Florida, Tampa, United States

A Practical Distance Measurement Improvement Technique for a SFCW-based Health ;; Monitoring Radar

M. Mercuri¹, P. Soh¹, D. Schreurs¹, P. Leroux², ¹KU Leuven, Heverlee (Leuven), Belgium, ²Thomas More Kempen, Geel, Belgium

Under-Sampled Measurement Technique for Hybrid-Filter-Bank Characterization in Real-Time 325 Applications

J. P. Magalhães², D. Ribeiro², J. M. Vieira¹, R. Gómez-García³, N. B. Carvalho², ¹Universidade de Aveiro, Aveiro, Portugal, ²Universidade de Aveiro / Instituto de Telecomunicações, Aveiro, Portugal, ³University of Alcalá, Alcalá de Henares, Spain

Spacial Vector Measurement Based on Six-Port Concept 329

V. Adler, K. Hoffmann, Czech Technical University in Prague , Prague 6, Czech Republic

Calibration for Noise Parameters Measurement with a VNA and without a Hot Noise Source 332

D. Pasquet¹, Z. Huang², C. Qiu², D. Lesenechal¹, P. Descamps¹, ¹LaMIPS, Caen, France, ²ENSEA, Cergy, France

Uncertainties in Complex Permittivity Extraction from Coplanar Waveguide Scattering- 335 Parameter Data

U. Arz¹, D. F. Williams², ¹PTB, Braunschweig, Germany, ²NIST, Boulder, United States

Microwave Interferometric Method for Metal Sheet Thickness Measurement 33;

K. Hoffmann, Z. Skvor, Czech Technical University in Prague , Prague 6, Czech Republic

Using a Fiber Optic Link to Extend Network Analysis Measurement for Spatially Dispersed 344 Components

L. Hausman, L. Oliva, MITEQ, Inc., Hauppauge, United States

Efficient Noise Extraction Algorithm and Wideband Noise Measurement System from 0.3 GHz to 67 GHz 347

H. V. Nguyen, N. Misljenovic, B. Hosein, Focus Microwaves Inc., Dollard-des-Ormeaux, Canada

2D Multisine Mapping for Robust 2 Band PA Modeling and 2D Predistorter Extraction 349

P. Roblin¹, N. Naraharisetti¹, C. Quindroit¹, S. Gheitanchi², V. Mauer², M. Fitton³, ¹Ohio State University, Columbus, United States, ²Altera Europe Limited, High Wycombe, United Kingdom, ³Altera, San Jose, United States

The Effects of Power Amplifier Nonlinearities and CFR on 64QAM HSDPA Waveforms 352

R. N. Braithwaite, Powerwave Technologies, Santa Ana, United States

An Automated System for Testing an Avionics Radio 355

T. O. Shafer, Spectralux Corporation, Redmond, United States

Intermodulation Distortion of Actuated MEMS Capacitive Switches 35:

X. Luo¹, Y. Ning¹, D. Molinero¹, C. Palego¹, J. C. Hwang¹, C. L. Goldsmith², ¹Lehigh University, Bethlehem, United States, ²MEMtronics Corporation, Richardson, United States

Residual Errors in Coplanar On-Chip 1-Port Calibration Caused by Standard Deviations 163

S. Berndt¹, T. Mrosk¹, F. Lenk¹, R. Doerner², ¹Hochschule Lausitz (FH), Senftenberg, Germany, ²Ferdinand-Braun-Institut, Berlin, Germany

Operation and Calibration of VNA-based Large Signal RF I-V Waveform Measurements System without using a Harmonic Phase Reference Standard 366

A. S. Aldoumani¹, P. J. Tasker¹, R. S. Saini², T. Williams², J. Lees¹, J. W. Bell¹, ¹Cardiff University, Cardiff, United Kingdom, ²Mesuro Limited, Pen coed, United Kingdom

A Novel Active Load-pull System with Multi-Band Capabilities 36:

S. Gustafsson, M. Thorsell, C. Fager, Chalmers University of Technology, Gothenburg, Sweden

Measurement Uncertainties due to Non-Ideal Calibration Standards for Unknown Thru Calibration 174

F. Lenk¹, R. Doerner², P. Kurpas², A. Rumiantsev³, ¹Hochschule Lausitz (FH), Senftenberg, Germany, ²Ferdinand-Braun-Institut, Leibniz-Institut für Höchstfrequenztechnik, Berlin, Germany, ³Brandenburg University of Technology Cottbus, Cottbus, Germany

On-Wafer Multi-Port Circuits Charaterization Technique with a Two-Port VNA 178

S. Quintanel¹, D. Pasquet², E. Bourdel¹, C. Duperrier¹, D. Lesenechal², T. Dinh², P. Descamps², ¹ENSEA, Cergy, France, ²LaMIPS, Caen, France

Using Lines of Arbitrary Impedance as Standards on the TRL Calibration Technique 382

J. Reynoso-Hernández¹, M. A. Pulido-Gaytán¹, A. Zárate-de Landa¹, J. E. Zúñiga-Juárez¹, J. R. Monjardín-López¹, A. García-Osorio², D. Orozco-Navarro², J. R. Loo-Yau², M. D. Maya-Sánchez¹, ¹CICESE, Ensenada, Mexico, ²CINVESTAV Guadalajara, Zapopan, Mexico

Development and Traceable Measurements of 0.800 mm Coaxial Airline: Broadband Standard to 186 GHz

T. Roberts, Y. Lee, J. Martens, B. Oldfield, Anritsu, Morgan Hill, United States

Removal of Measurement Artifacts Present in High-Power RF Transistor Loadpull Test-Fixtures 18:

T. A. Barbieri, P. H. Aaen, B. Noori, Freescale Semiconductor, Tempe, United States

Thermoelectric Reference Standards and a Measurement System for Microcalorimeters 195

J. Kwon, T. Kang, D. Lee, N. Kang, Korea Research Institute of Standards and Science (KRISS), Daejeon, Republic of Korea

IMS/ARFTG SESSION TH2D: Measurements Supporting Active Device Modeling*Chair: Nuno Borges Carvalho, Universidade de Aveiro**Co-Chair: John T. Barr, Agilent Technologies***Non-Intrusive Characterization of Active Device Interactions in High-Efficiency Power Amplifiers N/A**R. Hou¹, M. Spirito¹, J. Gajadharsing², L. C. de Vreede¹, ¹Delft University of Technology, Delft, Netherlands, ²NXP Semiconductors, Nijmegen, Netherlands**Synchronous Frequency Domain Measurements for the Extraction of X-parameters in Digital to Analog Transmitters N/A**

D. C. Ribeiro, P. M. Cruz, N. B. Carvalho, Instituto de Telecomunicacoes (IT), Aveiro, Portugal

A Precision Millimeter-Wave Modulated-Signal Source N/A

K. A. Remley, P. D. Hale, D. F. Williams, C. Wang, National Institute of Standards and Technology (NIST), Boulder, United States

Extremely Low-Frequency Measurements Using an Active Bias Tee N/AA. Nalli¹, A. Raffo¹, G. Avolio², V. Vadalà¹, G. Bosi¹, D. M. Schreurs², G. Vannini¹, ¹University of Ferrara, Ferrara, Italy, ²KU Leuven, Leuven, Belgium**IMS/ARFTG SESSION TH3D: Innovative Measurements Across the Spectrum (RF to THz)***Chair: David Blackham, Agilent Technologies**Co-Chair: Michael Janezic, National Institute of Standards and Technology (NIST)***Space-Resolved Measurement and 2D-Mapping of Material Parameters Using Multi-Angle Reflection Ellipsometry in W-Band N/A**

A. Cenanovic, L. P. Schmidt, University of Erlangen-Nuremberg, Erlangen, Germany

Accurate Permittivity Measurements with a Coaxial Resonator Independently of Coupling Level N/A

B. Garcia-Banos, A. J. Canos, J. M. Catala-Civera, P. J. Plaza-Gonzalez, Polytechnic University of Valencia, Valencia, Spain

Terahertz 7-port Reflectometer for S-parameter Measurements N/A

J. Yao, M. C. Wah, A*STAR, Singapore, Singapore

Dual Tone Approach for Unambiguous Six-Port Based Interferometric Distance Measurements N/A

S. Lindner, G. Vinci, F. Barbon, S. Mann, R. Weigel, A. Koelpin, University of Erlangen-Nuremberg, Erlangen, Germany